



Republic of the Philippines  
Department of Health  
**MARGOSATUBIG REGIONAL HOSPITAL**  
Margosatubig, Zamboanga del Sur



(0981) 141-4419

(062) 925-8033

✉ [mrh\\_gov@yahoo.com.ph/mrhgov@gmail.com](mailto:mrh_gov@yahoo.com.ph/mrhgov@gmail.com)

🌐 <http://mrh.doh.gov.ph>

Date: November 23, 2023

Validity: 60 days

Minimum Technical Requirements for  
**DIAGNOSTIC SET**

**PRODUCT SPECIFICATIONS**

- Ophthalmoscope
  - Six aperture selection dials
  - Triple sealed ophthalmoscope
  - Cross-linear polarizing red-free filter
  - Lens selection disc
    - At least 28 focusing lenses
- Otoscope
  - Fiber optics project cool light to the distal tip
  - No reflection or obstructions for view of the tympanic membrane
- Handle
  - 3.5 Volts dc
  - Lithium-ion battery
  - At least 100 minutes
  - Low battery indicator

Page | 1

**STANDARD ACCESSORIES:**

- Charger

**STANDARD REQUIREMENTS UPON SUBMISSION OF BIDDING DOCUMENTS**

- Soft copy of user manual

**REQUIREMENT/S IF AWARDED THE CONTRACT**

- Unit must be **BRAND NEW**.
- Manual in English Language
  - 2 copy of Operation manual (Hard copy)
  - Service manual with soft copy

PREPARED & REVIEWED BY

END USER: JESSICA B. OLIVEROS, RN, MN  
DESIGNATION: NURSE V

JAMES CARLTY D. LABADAN, ECE  
CHAIRPERSON

KEVIN S. CATAGUE, ECE  
VICE-CHAIRPERSON

JOEL L. LUNA  
MEMBER

IVAN PAUL L. TANO, ECE  
MEMBER

CRISABELLE G. GUERRA, MPA  
MEMBER

APPROVED BY:

RICHARD B. SISON, MD, MHA, FPCHA, FPSMS  
Medical Center Chief II



Republic of the Philippines  
 Department of Health  
**MARGOSATUBIG REGIONAL HOSPITAL**  
 Margosatubig, Zamboanga del Sur



(0981) 141-4419

(062) 925-8033

mrh\_gov@yahoo.com.ph/mrhgov@gmail.com

http://mrh.doh.gov.ph

Date: November 23, 2023

Validity: 60 days

Minimum Technical Requirements for  
**DIAGNOSTIC SET**

- Training Program
  - for end-user on equipment's operation with certificate
  - Technical training with certificate (When necessary)
- Three (3) years warranty on parts and service after commissioning and acceptance
- Certification on the availability of spare parts for the next 5 years
- Down time of machine will be added as extension of the warranty.
- With calibration certificate (When necessary)
- Supplier must respond to a service call before 72 hours and should be onsite not more than 7 days.
- **Section V. Special Conditions of the Contract**
  - Delivery Place: Margosatubig Regional Hospital, Zamboanga del Sur
- **Section VI Schedule of Requirements**
  - Completion/Delivery Period: \_\_\_\_ calendar days upon receipt of the Notice to Proceed including the delivery, installation, testing and commissioning, Partial delivery allowed within the completion/delivery period.

Page | 2

PREPARED & REVIEWED BY:

END USER: JESSICA A. OLIVEROS, RN, MN  
 DESIGNATION: NURSE

JAMES CARLTY D. LABADAN, ECE  
 CHAIRPERSON

KEVIN S. CORTAGUE, ECE  
 VICE-CHAIRPERSON

JOR L. LUNA  
 MEMBER

IVAN PAUL L. TANO, ECE  
 MEMBER

CRIS ABEL M. SUEVARRA, MPA  
 MEMBER

APPROVED BY:

RICHARD B. SISON, MD, MHA, FPCHA, FPSMS  
 Medical Center Chief II