



May 12, 2022

Validity: 60 days

Minimum Technical Requirements for
Oscillating saw handpiece

PRODUCT SPECIFICATIONS

- Brushless DC motor
 - 12000 cycles per minute
- Power supply
 - Ni-MH Battery : 9.6V , 2100 mAh
- Material
 - Stainless steel 304
- Autoclaveable

STANDARD ACCESSORIES:

- Battery
- Battery box
- Charger
- Transfer shield
- Blade

STANDARD REQUIREMENTS UPON SUBMISSION OF BIDDING DOCUMENTS

- Soft copy of user manual

REQUIREMENT/S IF AWARDED THE CONTRACT

- Manual in English Language
 - 2 copy of Operation manual (Hard copy)
 - Service manual with soft copy
- Training Program
 - for end-user on equipment's operation with certificate
 - Technical training with certificate (When necessary)
- Two (2) years warranty on parts and service after commissioning and acceptance
- Certification on the availability of spare parts for the next 5 years
- With calibration certificate (When necessary)
- **Section V. Special Conditions of the Contract**
 - Delivery Place: Margosatubig Regional Hospital, Zamboanga del Sur



Republic of the Philippines
Department of Health
MARGOSATUBIG REGIONAL HOSPITAL
Margosatubig, Zamboanga del Sur



Contact No. (0917) 715-0041 / e-mail add: mrh_gov@yahoo.com.ph/mrhgov@gmail.com website:<http://mrh.doh.gov.ph>

➤ **Section VI Schedule of Requirements**

- Completion/Delivery Period: 120 calendar days upon receipt of the Notice to Proceed including the delivery, installation, testing and commissioning, Partial delivery allowed within the completion/delivery period.

PREPARED & REVIEWED BY:

MICHAEL D. SALVADOR, RN, MAN, DMngt
Nurse IV

JAMES CARLY D. LABADAN, ECE
Chairman

JOEL L. LUNA
Vice-Chairman

IVAN PAUL L. TANO, ECE
Member

CRIS ABEL M. GUEVARRA, MPA
Member

KEVIN S. CATAGUE, ECE
Member

ILLUMAR PANES, RN
Member

APPROVED BY:

RICHARD B. SISON, MD, MHA, FPCHA, FPSMS
Medical Center Chief II